



IgG

IgG (Immunoglobulin G)

Turbidimetry

Quantitative determination of human immunoglobulin G (IgG) IVD

Store 2 - 8°C.

INTENDED USE

The IgG is a quantitative turbidimetric test for the measurement of IgG in human serum or plasma.

PRINCIPLE OF THE METHOD

Anti-human IgG antibodies when mixed with samples containing IgG, form insoluble complexes. These complexes cause an absorbance change, dependent upon the IgG concentration of the patient sample, that can be quantified by comparison from a calibrator of known IgG concentration.

CLINICAL SIGNIFICANCE

IgG is the most important immunoglobulin produced by plasma cells, and represents about 75% of the total immunoglobulins. Its main function is to neutralize toxins in tissue spaces.

IgG deficit may be due to a congenital primary disturbance (immunodeficiency congenital and acquired) and is a special risk in children.

Polyclonal hyperimmunoglobulinemia is the normal response to infections, especially in hepatitis and cirrhosis as well as autoimmune diseases.

Increases of monoclonal IgG are found in multiple myeloma, lymphocytic leukemia, and Waldenström macroglobulinemia.

REAGENTS

Diluent (R1)	Tris buffer 20 mmol/L, PEG 8000, pH 8.3. Sodium azide 0.95 g/L.
Antibody (R2)	Goat serum, anti-human IgG, pH 7.5. Sodium azide 0.95 g/L.
Optional	Cod: 1102003 PROT CAL.

CALIBRATION

The assay is calibrated to the Reference Material CRM 470/RPPHS (Institute for Reference Materials and Measurements). It must be used the PROT CAL Calibrator to calibrate the reagent. The reagent (both monoreagent and bireagent) should be recalibrated every month, when the controls are out of specifications, and when changing the reagent lot or the instrument settings.

PREPARATION

Reagents: Ready to use.

Calibration Curve: Prepare the following PROT CAL Calibrator dilutions in NaCl 9 g/L as diluent. Multiply the concentration of the IgG calibrator by the corresponding factor stated in table below to obtain the IgG concentration of each dilution.

Calibrator dilution	1	2	3	4	5	6
Calibrator (μL)	--	10	25	50	75	100
NaCl 9 g/L (μL)	100	90	75	50	25	-
Factor	0	0.1	0.25	0.5	0.75	1.0

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C and contaminations are prevented during their use. Do not use reagents over the expiration date.

Reagent deterioration: The presence of particles and turbidity.

Do not freeze; frozen Antibody or Diluent could change the functionality of the test.

ADDITIONAL EQUIPMENT

- Thermostatic bath at 37°C.
- Spectrophotometer or photometer thermostatable at 37°C with a 600 nm filter (580 - 620 nm).

SAMPLES

Fresh serum or plasma. EDTA or heparin should be used as anticoagulant. Stable 7 days at 2-8°C or 3 months at -20°C.

The samples with presence of fibrin should be centrifuged.

Do not use highly hemolyzed or lipemic samples.

PROCEDURE

1. Bring the reagents and the photometer (cuvette holder) to 37°C.

2. Assay conditions:

Wavelength : 600

Temperature : 37 °C

Cuvette length path : 1cm

3. Adjust the instrument to zero with distilled water.

4. Pipette into a cuvette:

Reagent R1	800 μL
Sample or Calibrator	10 μL

5. Mix and read the absorbance (A_2) after 5 minutes.

Reagent R2	200 μL
------------	--------

6. Mix and read the absorbance (A_2) of calibrators and sample exactly 5 minutes after the R2 addition.

Audit Diagnostics has instruction sheets for several automatic analyzers. Instructions for many of them are available on request.

CALCULATIONS

Calculate the absorbance difference ($A_2 - A_1$) of each point of the calibration curve and plot the values obtained against the IgG concentration of each calibrator dilution. IgG concentration in the sample is calculated by interpolation of its ($A_2 - A_1$) in the calibration curve.

QUALITY CONTROL

Control sera are recommended to monitor the performance of manual and automated assay procedures. Audit Diagnostics PROT CONTROL.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES⁴

Between 700 - 1600 mg/dL. Each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

1. Measurement range: Up to 2000 mg/dL under the described assay conditions. Samples with higher concentrations, should be diluted 1/5 in NaCl 9 g/L and re-tested again. The linearity limit and measurement range depends on the sample to reagent / ratio. It will be higher by decreasing the sample volume, although the sensitivity of the test will be proportionally decreased.
2. Detection Limit: Values less than 180 mg/dL give non-reproducible results.
3. Prozone effect: No prozone effect was detected upon 8000 mg/dL.
4. Sensitivity: $\Delta 0.6$ mA. mg/dL at 359 mg/dL.
5. Precision: The reagent has been tested for 20 days, using three levels of serum in a EP5-based study.

EP5	CV (%)		
	340.3 mg/dl	801.96 mg/dl	1517.5 mg/dl
Total	2.1%	2.8%	4.8%
Within Run	0.9%	0.7%	1%
Between Run	1.5%	1.5%	1.8%
Between Day	1%	2.2%	4.4%

6. Accuracy: Results obtained using this reagent (y) were compared to those obtained using the Elecsys method from Roche. 79 samples ranging from 450 to 2600 mg/dL of IgG were assayed. The correlation coefficient (r) was 0.94 and the regression equation $y = 0.957x + 105.67$.

The results of the performance characteristics depend on the used analyzer.

INTERFERENCES⁵⁻⁶

Hemoglobin (10 g/L), bilirubin (20 mg/dL) and lipemia (10 g/L), do not interfere. Rheumatoid factors may interfere at 300 IU/mL. Other substances may interfere^{5,7}.

NOTES

1. Clinical diagnosis should not be made on findings of a single test result, but should integrate both clinical and laboratory data.

BIBLIOGRAPHY

1. Clinical Guide to Laboratory Tests, Edited by NW Tietz W B Saunders Co., Philadelphia, 483, 1983.
2. Skoug Jonh W et al. Clin Chem 1988; 34/2: 309 - 315
3. Pesce AJ and Kaplan, LA. Methods in Clinical Chemistry. The CV Mosby Company, St. Louis MO, 1987.
4. Dati F et al. Eur J Clin Chem Clin Biochem 1966; 14: 401-406.
5. Young DS. Effects of disease on clinical laboratory tests, 3th ed. AACC Pres, 1997
6. Friedman and Young. Effects of disease on clinical laboratory tests, 3th ed. AACC Pres, 1997.