



LACTATE Liquid Reagent

INTENDED USE:

For the quantitative determination of lactate in plasma.


SUMMARY AND EXPLANATION:

Aerobic glycolysis markedly increases blood lactate and causes some increase in pyruvate levels, especially with prolonged exercise. The common cause for increased blood lactate and pyruvate is anoxia resulting from such conditions as shock, pneumonia and congestive heart failure. Lactic acidosis may also occur in renal failure and leukaemia. Thiamine deficiency and diabetic ketoacidosis are associated with increased levels of lactate and pyruvate.

Lactate measurements that evaluate the acid-base status are used in the diagnosis and treatment of lactic acidosis (abnormally high acidity in the blood).

PRINCIPLE OF THE TEST:

Lactate-oxidase (LOX) cleaves lactate into pyruvate and H₂O₂, which reacts in the presence of peroxidase (POD) with 4-aminopyridine and phenol to a red chromogen dye. The increase in absorbance due to the dye formation is proportional to the lactate concentration in the sample.



WARNINGS AND PRECAUTIONS:

For In Vitro Diagnostics Use Only - For Professional Use Only

Carefully read instructions for use. Deviations from this procedure may alter performance of the assay.

Colour and Appearance: Clear, slightly pink liquid.

Any significant changes from the above could indicate that the assay might be compromised. Refer to Laboratory's QC program for actions to be taken. In case of serious damage to the bottle and/or cap, resulting in product leakage and/or contamination, do not use the reagent pack and contact your distributor.

Safety precautions: **Caution:** Take all necessary precautions required when handling laboratory reagents. Material Safety Data Sheet is available upon request.

Handling precautions:

- Do not use components past the expiry date stated on the Bottles.
- Do not Freeze Reagents.
- Do not use components for any purpose other than described in the "Intended Use" section.
- Do not interchange caps among components as contamination may occur and compromise test results.
- Refer to local legal requirements for safe waste disposal.

COMPONENT COMPOSITION:

Component	Ingredients	Concentration in Tests
Reagent 1	MONOREAGENT	
	TRIS-Buffer pH 7.5	50.0 mmol/l
	LOX	0.2 KU/l
	POD	3 KU/l
	4-aminopyridine	0.4 mmol/l

REAGENT PREPARATION AND STABILITY:

Reagent 1 is ready for use.
Before use, mix reagent by gently inverting each bottle.

The reagents are stable up to the expiry date on the label when stored at +2°C to +8°C.
On-board stability: The reagent will be stable for up to 28 days

Coloration of the reagent (reagent blank at 505 nm - 546 nm, 1 cm > 0.2) indicates contamination or damage by storage at higher temperatures.

TYPE OF SPECIMEN:

Use heparin or fluoride plasma as specimen

Large and Variable change in lactate concentration may occur after sample collection. These changes may be minimised by using fluoride-oxalate as anti-coagulant, keeping samples on ice and separating plasma from blood without delay.

While lactate levels are stable in whole blood, the separated plasma may be stored at 2-8°C for 48 hours prior to use. Plasma lactate is stable up to at least 1 month stored frozen at -20°C

Sample preparation

- Draw blood with a minimum of stasis from fasting, resting patients into tubes containing fluoride/oxalate as anticoagulant.
- Mix well by gentle inversion at least 6 times. Cool blood in an ice bath.
- Within 30 minutes separate the plasma from blood by centrifugation at 400 x g for 10 minutes.
- Avoid excessive forces which contribute haemolysis. Note: If separation of plasma from blood (kept on ice) is delayed, lactate values may variably increase by about 10% after 1 hour, and up to 15% after 2 hours

TEST PROCEDURE:

Assay procedure

Refer to relevant user's manual for instructions on instrument start-up, loading components and samples calibration, sample testing procedures, calculating and reporting results.

Calibration

- Using recommended Calibrator, calibrate the assay.
- When using a new reagent kit or changing lot number.
- Following preventive maintenance or replacement of a critical part.
- When Quality Controls are out of range.

Quality Control

All clinical laboratories should establish an Internal Quality Control program. Verify instrument and reagent performance with recommended controls or similar. The values obtained for QC should fall within manufacturer's acceptable ranges or should be established according to the Laboratory's QC Programme.

Controls should be assayed:

- Prior to reporting patient results.
- Following any maintenance procedure.
- At intervals established by the Laboratory's QC Programme.

CALCULATION:

The analyser automatically calculates the lactate concentration in the sample.
Conversion factor: mg/dl = Qty in mmol/l x 9.015.

EXPECTED VALUES:

	mg/dl	mmol/l
Plasma	4.5 - 19.8	0.5 - 2.2

PERFORMANCE CHARACTERISTICS:

Performance results can vary with the instrument used. Data obtained in each individual laboratory may differ from these values.

Linearity

Linear up to 140mg/dl.

For samples with a higher concentration, dilute 1:1 with 0.9% NaCl (9g/l) and re-assay. Multiply result by 2.

Interfering substances: Results of study are as follows:

Bilirubin: < 10% interference up to 400mg/L Bilirubin.
Haemolysis: < 10% interference up to 20g/L Haemoglobin.
Lipemia: <10% interference up to 10g/L Intralipid

Sensitivity: The Lowest Detectable Level was estimated at 2.0 mg/dl (0.2 mmol/l).

Precision:

Within Run				Between Run			
Level	Mean mg/dl	SD mg/dl	% CV	Level	Mean mg/dl	SD mg/dl	% CV
Level 1	13.6	0.25	1.80	Level 1	14.3	0.40	2.79
Level 2	45.7	0.46	1.01	Level 2	46.8	0.38	0.82

Method Comparison:

A method comparison between this Lactate test (y) and another commercially available test (x), gave the following results:

y=1.087x-1.4738	R=0.9992	Sample range: 8.3-55 mg/dl
-----------------	----------	----------------------------

BIBLIOGRAPHY:

- Bablok W et al. A general regression procedure for method transformation. J Clin Chem Biochem. 1988;26:783-790.
- Barhan D, Trinder P. An improved colour reagent for the determination of blood glucose by the oxidase system. Analyst. 97:1972-142.
- Bergmeyer HU (ed). New York, NY: Academic Press Inc; 1974:1464.
- Burtis CA, Ashwood ER, eds. Tietz Textbook of Clinical Chemistry, 2nd ed. Philadelphia, PA: W.B. Saunders Co; 1994:576.
- Gutmann I, Wahlefeld A. Methods of Enzymatic Analysis, 2nd ed.
- Noll F. Methods of Enzymatic Analysis, 2nd ed. Bergmeyer HU ed.
- Passing H, Bablok W. A new biometrical procedure for testing the equality of measurements from two different analytical methods. J Clin Chem Clin Biochem. 1983;21:709-720.
- Tietz NW. clinical Guide to Laboratory Tests, 3rd ed. Philadelphia, PA: WB Saunders Co; 1995: 382 - 383
- Trinder P. Determination of glucose in blood using glucose oxidase with an alternative oxygen acceptor. Ann. Clin. Biochem. 1969;6:24.